



**Measured values:**

CWL = 5233,69 nm...OK  
 HW = 124,76 nm...OK  
 Cuton 5 % = 5118,85 nm  
 Cutoff 5 % = 5347,67 nm  
 HP1 = 5171,31 nm  
 HP2 = 5296,06 nm  
 Tpeak = 73,74 %  
 Tavg = 71,01 %  
 Slope 1 = 1,29 %  
 Slope 2 = 1,26 %

**Transmittance:**

T high min 5195 — 5271 nm = 73,74 %...OK

# SPECTROGON

Optical filters • Coatings • Gratings

Date: 2023-02-01

Operator: FB

**Article:**

Art no: 713M

NB-5250-125 nm Ø25.4x1.0 mm

**Specifications:**

CWL: 5250 ±40 nm

HW: 125 ±25nm

Blocking: 100 — 30000 nm (0,1 %)

T high min 60 %

**Measurement conditions:**

Temperature: 296 K AOI: 0°

**Spectrogon:**

T nr: 69207

Calibration traceable to NIST.  
 Spectrogon products conform in all aspects with the RoHS Directive.

# Kokyo

株式会社 光響

Email : [info@symphotony.com](mailto:info@symphotony.com)  
 Web : <https://www.symphotony.com/>